

**NEW**



Quality and reliability is our tradition

**KYORITSU**

# AC/DC DIGITAL CLAMP METER KEW SNAP 2009R

**CAT.IV 600V**

**TRUE RMS**

**φ55mm**

**TRUE  
RMS**

**CAT.IV  
600V**

**φ55**

**MAX  
2000A**

**DC  
AC A**

**DC  
AC V**

**Ω**

**Continuity**

**Hz**

**PEAK  
10ms**

**AVG**

**OUT  
PUT**

- **Designed to meet international safety standard IEC 61010-1 CAT. IV 600V.**
- **CAT.IV Clamp Meter can measure the Voltage and Current in both very low and high power circuits.**
- **Provides a wide measuring range from 0 up to 2000A AC/DC.**
- **Accurate True-RMS reading of AC current or voltage with distorted waveform.**
- **Measures current and voltage variation as short as 10 msec with peak-hold feature.**
- **AVERAGE function to allow for easy reading of fluctuating current or voltage.**
- **Provides output to a recorder for current variation recording.**

**Actual Size**

**KYORITSU ELECTRICAL INSTRUMENTS WORKS,LTD.**

<http://www.kew-ltd.co.jp>

● Specifications

KEW 2009R	
AC A	400.0/2000A ±1.3%rdg±3dgt(0~400A, 150~1700A)(45~66Hz) ±2.0%rdg±5dgt(0~400A, 150~1700A)(20Hz~1kHz) ±2.3%rdg±3dgt(1701~2000A)(45~66Hz)
DC A	400.0/2000A ±1.3%rdg±2dgt
AC V	40.00/400.0/750V ±1.0%rdg±3dgt(45~66Hz) ±1.5%rdg±5dgt(20Hz~1kHz)
DC V	40.00/400.0/1000V ±1.0%rdg±2dgt
Ω	400.0/4000Ω ±1.5%rdg±2dgt (Electronic protection (no fuse) up to 600V)
Continuity buzzer	Buzzer sounds below 20Ω
Hz	10.0~4000Hz ±1.5%rdg±5dgt (Input sensitivity Current:more than 40A Voltage:more than 10V)
Output	Recorder : DC400mV against AC/DC400.0A DC200mV against AC/DC2000A
Withstand voltage	6880V AC for 5 seconds
Conductor size	φ55mm max.
Applicable standards	IEC 61010-1 CAT.IV 600V CAT.III 1000V IEC 61010-031, IEC 61010-2-032 IEC 61326-1, IEC 61326-2-1
Power source	R6(1.5V)x2
Dimensions/Weight	250(L) × 105(W) × 49(D)mm / Approx. 540g(including batteries)
Accessories	7107(Test leads) 8201(Output plug) 9094(Carrying case) R6(AA)(1.5V) × 2, Instruction manual
Optional	8008(Multi-tran)(AC only) 7014(Output cord)

● Accessories



● Optional



● Selection Guide

	2009R	2003A	2002R	2002PA
AC A	2000A	2000A	2000A	2000A
AC V	750V	750V	750V	750V
DC A	2000A	2000A	—	—
DC V	1000V	1000V	1000V	1000V
Ω	4000Ω	4000Ω	400kΩ	400kΩ
Continuity buzzer	●	●	●	●
Frequency	4000Hz	—	—	—
Data hold	●	●	●	●
Peak hold	●	—	●	●
MAX hold	—	●	—	—
Average	●	—	—	—
DC OUTPUT	●	●	●	●
True RMS	●	—	●	—
IEC 61010-1	CAT.IV 600V	CAT.IV 600V	CAT.III 600V	CAT.III 600V



**Safety Warnings :**

Please read the "Safety Warnings" in the instruction manual supplied with the instrument thoroughly and completely for correct use. Failure to follow the safety rules can cause fire, trouble, electrical shock, etc. Therefore, make sure to operate the instrument on a correct power supply and voltage rating marked on each instrument.

■ For inquires or orders :



**KYORITSU ELECTRICAL INSTRUMENTS WORKS, LTD.**

No.5-20,Nakane 2-chome, Meguro-ku, Tokyo, 152-0031 Japan  
Phone:81-3-3723-0131  
Fax:81-3-3723-0152  
E-mail:info@kew-ltd.co.jp  
Factory:Ehime, Japan



<http://www.kew-ltd.co.jp>

ISO9001:2000 APPROVED BY BUREAU VERITAS Certification



In consideration of the environment, soy ink and recycled paper were used in this publication.

● The contents of this leaflet are subject to change without notice.

KEW SNAP 2009R-1E Jun. 10 AD